



## Case Report

# Unusual leukoagglutination: A rare haematological finding



Massimo Daves <sup>ID a</sup>, Vincenzo Roccaforte <sup>ID b,c,d,\*</sup>, Erika Jani <sup>a</sup>, Evi Peer <sup>a</sup>

<sup>a</sup> Hospital of Bolzano, Bolzano, Italy

<sup>b</sup> ASST Grande Ospedale Metropolitano Niguarda, Milan, Italy

<sup>c</sup> University of Study, Milan, Italy

<sup>d</sup> Hospital Alessandro Manzoni, Lecco, Italy

## ARTICLE INFO

### Article history:

Received 24 April 2020

Accepted 30 November 2020

Available online 21 December 2020

## Introduction

The complete blood cell count (CBC) is a crucial test for diagnosing and managing of a variety of hematological disorders, provided that a high degree of quality can be assured throughout the testing process. Although the modern hematology analyzers are now capable to generate rapid and accurate results in most clinical circumstances, spurious results related either to white blood cells (WBC) or other parameters of CBC may occur, then leading to further unnecessary investigations and/or inappropriate clinical management.<sup>1</sup> Several causes of spuriously high and low WBC counts have been reported, thus including platelets aggregate and large platelets, presence of nucleated red blood cells, red blood cells resistant to lysis, cryoglobulins, cryofibrinogen, lipidic droplets, microorganism, embolism of adipose tissue. Since WBC aggregates have been occasionally reported as causes of spuriously low test results,<sup>1</sup>

we describe here the case of spurious WBC count observed with the hematology analyzer Sysmex XN 9000.

## Case report

In September 2019 a 70-year-old male individual underwent routine blood testing before elective surgical procedure for abdominal hernia. The examinations included CBC, routine coagulation tests (prothrombin time and activated partial thromboplastin time), and clinical chemistry (electrolytes, creatinine, glucose, total protein, urea, C-reactive protein, GAMMA-GLUTAMYLTRANSFERASE, ALANINE AMINOTRANSFERASE, ASPARTATE AMINOTRANSFERASE, TOTAL BILIRUBIN AND LACTATE DEHYDROGENASE). With the exception of gamma-glutamyltransferase, which was slightly elevated, all other biochemical parameters were within the reference ranges.

The past medical history revealed that the patient underwent neurosurgery for meningioma in 2010 and ears, nose, and throat (ENT) surgery for neoformation of vocal cords in 2012. He was also diagnosed as having liver cirrhosis,

\* Corresponding author at: Clinical Pathology Laboratory, Hospital Alessandro Manzoni, Lecco, Italy.


E-mail address: [v.roccaforte@asst-lecco.it](mailto:v.roccaforte@asst-lecco.it) (V. Roccaforte).

<https://doi.org/10.1016/j.htct.2020.11.003>


alcohol-related (Child-Pugh Classification A), along with arterial hypertension, and arthrosis.

For CBC blood had been collected with a 21-gauge needle into a blood tube containing dipotassium ethylenediaminetetraacetate (K2EDTA) (Vacutte, Greiner Bio-One International®). Within 30 min after blood collection, the CBC was performed on a Sysmex XN analyzer (Sysmex®, Kobe, Japan). The hematology analyzer XN enumerates and classifies blood cells with a combination of direct current detection, by using flow cytometry technology. More specifically, cell enumeration is based on semiconductor laser counting, and cells are then classified by irradiation with a 633 nm laser beam and analysis of forward scattered light, side scattered light and side fluorescent light. The three signals are integrated to differentiate and count WBC, nucleated red blood cells (NRBC), reticulocytes and platelets, as well as for detecting abnormal and immature cells.<sup>2</sup> The performance of the Sysmex XN-series has been extensively evaluated by Seo et al., as described elsewhere.<sup>3</sup> All parameters of CBC were within the reference ranges except for a minimal leukocytosis of  $10.9 \times 10^9/L$  (reference range,  $3.6\text{--}10.5 \times 10^9/L$ ) with a differential leukocyte count of 57.2% neutrophils ( $6.3 \times 10^9/L$ ), 35.2% lymphocytes ( $3.8 \times 10^9/L$ ), 7.1% monocytes ( $0.8 \times 10^9/L$ ), 0.3% eosinophils and 0.2% basophils. Nevertheless, an analyzer flag accompanied the data, suggestive for the presence of immature granulocytes (alarm message: “IG present”, with a count of 5.4% of immature granulocytes). The “atypical” cloud observed in the WDF (White Cell Differential) scattergram was responsible for the generation of the alert “IG presence”. Due to the presence of the flag “IG presence” a blood film was prepared. The microscopic observation of the blood film showed normal erythrocyte and platelet morphology, and no immature leukocyte forms were seen. Nevertheless, some leukocyte aggregates, mostly composed of neutrophils, were detected. Microscopic observation at 400 magnifications revealed the presence of one or two aggregates in about 50% of the 100 observed fields. These aggregates were found to be composed of a variable number of neutrophils (from 2 to 5–10, in rare cases even more) (Fig. 1), some of which containing also one or two monocytes. The scattergram patterns provided by the XN analyzer are shown in Fig. 2. Interestingly, in the area corresponding to the neutrophil region in the scattergram provided by the WDF channel, a small cloud was observed that could have been generated by cells with a nucleic acid content apparently higher and which was interpreted as being small households. In the present case, this area may have been produced by little aggregates. Although the WBC count generated by the WNR channel is that finally reported by the analyzer, the count generated by the WDF channel can also be accessed through a dedicated menu. In our case report the two counts were similar ( $10.9$  and  $10$   $\text{WBC} \times 10^9/\mu\text{L}$  in the WNR and WDF channels, respectively).

The presence of these aggregates prompted us to investigate whether a potential underestimation of WBC count may have occurred, since the manufacturer claims that the analyzer may erroneously report a low white blood cell count when leukocyte aggregates are present.<sup>4</sup> Then, we performed another WBC count on the same sample using a flow cytometer (Navios Flow Cytometer® - Beckman Coulter, CA, USA) based on the principle of light scatter and intensity of fluores-



**Fig. 1 – Granulocytes aggregates on the blood smear.**



**Fig. 2 – Scattergram patterns of the WBC count in the WDF and WNR Channel. The arrow indicates the atypical cloud probably generated by some of the aggregates.**

cence signal (employing propidium iodide, a DNA intercalating agent, for cell count).<sup>7</sup> The fluorescent plot provided by the instrument showed unexpected fluctuations of fluorescence, which were positively correlated with the polyploid DNA content, corresponding to passage of leukocyte aggregates within the detection chamber with a WBC count of  $10,130/\mu\text{L}$ .

Although EDTA may induce leukocyte aggregation, this shall be considered a relatively rare finding compared to the well-known phenomenon of EDTA-induced thrombocytopenia.<sup>9</sup> The total WBC count in our citrate sample was found to be  $10.0 \times 10^9/\text{L}$ . Interestingly, the same type of leukocyte aggregates was observed in this sample by blood smear revision. We also repeated the CBC and blood smear revision after incubation at  $37^\circ\text{C}$  for 2 h and for 24 h without observing any significant difference for both the WBC count and the presence of aggregates. After 3 days the WBC count was still  $9.8 \times 10^9/\text{L}$  and identical aggregates were seen on the blood smear. The investigation was based on pre-existing samples after making the sample anonymous and in accordance with the Declaration of Helsinki and under the terms of all relevant local legislations.

## Discussion

The phenomenon of leukocyte agglutination in peripheral blood is reported as a rare and possible transient event.<sup>1</sup> The scientific literature describes some cases associated with cancers, severe infections, autoimmune disorders or hepatic diseases.<sup>8</sup> Although the phenomenon of EDTA-induced platelet agglutination is relatively frequent,<sup>9</sup> some cases of in vitro leukocyte aggregation have also been reported. Spurious WBC counts due EDTA-induced leukoagglutinins has been reported in association with cirrhosis, infections, autoimmune diseases, uremia, malignancies and immunosuppression.<sup>8</sup> Anand M et al. concluded that the leukocyte aggregation in vitro is a time-dependent phenomenon.<sup>10</sup> In our case the larger aggregates were overlooked by the analyzer, thus leading to underestimating the WBC count.<sup>1,4</sup> Instead, the scattergram generated by the WNR channel was normal. In WDF channel, WBC are counted and differentiated, whilst abnormal cells (e.g., immature WBC and atypical lymphocytes) are detected with a dedicated reagent. The WNR channel counts WBC and perform differential counting of basophils and NRBC with a different reagent.<sup>4</sup> In this case specimen was stored at room temperature for 24 h, confirming the presence of the leukocyte aggregates. Since the presence of cold agglutinins reacting with WBC antigens is a similar phenomenon,<sup>1</sup> we repeated the CBC and the blood film after incubation at 37 °C to rule out this possibility, but we failed to observe any difference. In a previous article, Tantanate described an incorrect WBC count using an XN analyzer, showing that the WBC count was very low in the WNR channel but was normal in the WDF channel as well as using a different analyzer and after blood smear revision. It was hence concluded that these differences were due to the use of different reagents in the two channels (i.e., more acid in the WNR channel compared to the WDF channel). The more acid reagent may cause the lysis of fragile WBC, especially in the presence of abnormal WBC or in patients receiving cytotoxic drugs.<sup>6</sup> The mechanism leading to leukocyte aggregation has not been entirely elucidated, though more frequently attributed to EDTA.<sup>1</sup> Unlike the case earlier published by Sultan and Irfan<sup>8</sup> who report on a pregnant woman with spurious low WBC count due to EDTA-induced agglutination (no leukocyte aggregates could be observed in a paired citrate-plasma sample), the leukocyte aggregates in our case were also present in the citrate sample.

Interestingly, Hoffman reported that EDTA-induced leukoagglutination may be resolved by adding kanamycin

sulphate to the sample but, as earlier discussed, agglutination was not probably caused by EDTA in our case.<sup>10</sup> Unfortunately, the patient did not show up for further evaluation, consequently we were unable to perform further evaluations as the collection of a blood sample use of Anticoagulant Citrate Dextrose Solution-A (ACD-A) that in some cases is preferred because this phenomenon cannot be always eliminated using citrate as an additive. Although we cannot provide a definitive explanation on the real nature of these aggregates, it is important that laboratory professionals acknowledge the presence of leukocyte aggregates by correctly interpreting tests results, as well as the appearance of abnormal patterns in the scattergrams and flag, as in this case the alert "IG" message, of their hematological analyzers.

## REFERENCES

- Zandecki M, Genevieve F, Gerard J, Godon A. Spurious counts and spurious results on haematology analysers: a review. Part II: white blood cells, red blood cells, hemoglobin, red cell indices and reticulocytes. *Int J Lab Hematol.* 2007;29:21–41.
- Daves M, Zagler EM, Cemin R, Gnech F, Joos A, Platzgummer S, et al. Sample stability for complete blood cell count using the Sysmex XN haematological analyser. *Blood Transfus.* 2015;13:576–82.
- Seo JY, Lee ST, Kim SH. Performance evaluation of the new hematology analyzer Sysmex XN-series. *Int J Lab Hematol.* 2015;37:155–64.
- Sysmex Automated Hematology Analyzer XN-Series (XN-1000) Instructions for use. Sysmex Corporation Code N.BD562985, Date of last Revision: February 2014.
- Beckman Coulter Navios Instructions for Use Manual. PN 773232AH. 2015 Beckman Coulter.
- Sultan S, Irfan SM. In vitro leukoagglutination: a rare hematological cause of spurious leukopenia. *Acta Med Iran.* 2017;55:408–10.
- Lippi G, Plebani M. EDTA-dependent pseudothrombocytopenia: further insights and recommendations for prevention of a clinically threatening artifact. *Clin Chem Lab Med.* 2012;50:1281–5.
- Anand M, Gulati HK, Joshi AR. Pseudoleukopenia due to ethylenediaminetetraacetate induced leukoagglutination in case of hypovolemic shock. *Indian J Crit Care Med.* 2012;16:113–4.
- Hoffmann JJ. EDTA-induced pseudo-neutropenia resolved with kanamycin. *Clin Lab Haematol.* 2001;23:193–6.